

University of Kalyani



Tender Ref. No.: KU/NIET/CHEM/FIST/2025-26/01

Tender ID: 2025_UOK_922430_1

CORRIGENDUM-1

Name of the Work: Purchase of Modular TCSPC FLUORESCENCE LIFETIME MEASUREMENT SYSTEM and UV-VIS NIR Spectrophotometer to the Department of Chemistry, University of Kalyani.

Sl No.	Matter of Corrigendum	Shown as per NIT	To be Read as
1	Item specification (MODULAR TCSPC FLUORESCENCE LIFETIME MEASUREMENT SYSTEM)		
1.1	Excitation Source (Sl No.1)	i. 405 – 410 nm Pulsed Laser Diode, < 100 ps	i. 405 – 410 nm Pulsed Laser Diode, < 100 ps, Repetition rate at least 20 MHz
1.2	Sensitivity (SNR) (Sl No. 10)	≥ 30000 : 1	This parameter will be excluded
1.3	Upgradability and flexibility (Sl No. 11)	The system must be upgradable with provision for multiple monochromators, flexibility for attachment of different light sources and L-format and T-format (optional) optical arrangements.	This parameter will be excluded

Bidders are advised to note the above before submission of their bids.

Sd/–
Tender Inviting Authority
University of Kalyani